

产品分析证书
Certificate of Analysis

中文名称: 高车前素

English Name: Hispidulin

别名 (Alias): Hispidulin; Methoxyapigenin

产品编码 (Cat. No.): BP0730


CAS Number: 1447-88-7

分子式 (M. F.): C₁₆H₁₂O₆

分子量 (M. W.): 300.266

批号 (Batch No.): PRF8030603

报告日期 (Report date): 2017年3月06日



检验结果 (Analytical result):

检验项目 (Test Item)	检验指标 (Specifications)	检验结果 (Results)
外观 Appearance	Light yellow powder	Light yellow powder
纯度 Purity (HPLC-DAD, 334nm)*	≥98.0%	98.22%
质谱 Mass	300.266±1	Conforms
核磁 NMR	Comply with the structure	Conforms
水分 Water Content	<3.0%	1.02%

* 色谱图见附件 (Please find HPLC chromatography attached.)

检测方法 (Test Method): Column: Thermo Hypersil GOLD AQ C18, 4.6*250mm, 5.0um; Column temperature: 30°C; Detection Mode: UV334nm; Flow

Rate: 1.0ml/min; Sample dissolution: Methanol; Mobile Phase: A: 0.1% Phosphate acid in water B: Acetonitrile; Gradient elution: B, 25%-45%, 20min.

贮存条件 (Storage): 2~8°C, protected from light, keep package airtight when not in use.

复测期 (Retest date): two years (2019-03-05) under conditions list above.

QC: Zhang Ling

Date: 2017年3月06日

QA: Wu Qi

Date: 2017年3月06日

备注 (Remarks): The sample solutions should be prepared and used on the day of receipt. It is the best preparing the solutions immediately before use. If the solutions have to be made up in advance, it should be made as aliquots in tightly sealed vials at less than -20°C. Generally, these might be useable for up to two weeks.
In case of quality issue, please contact us within 15 days after receipt of the product.


Tel: +86-28-82633397 Fax: +86-28-82633165

http://www.phytopurify.com Email: sales@biopurify.com biopurify@gmail.com

SAMPLE INFORMATION

Sample Name:	Hispidulin PRF8030603	Acquired By:	System
Sample Type:	Standard	Sample Set Name:	
Vial:	56	Acq. Method Set:	Hispidulin
Injection #:	1	Processing Method:	Samples
Injection Volume:	10.00 ul	Channel Name:	334.0nm
Run Time:	25.0 Minutes	Proc. Chnl. Descr.:	PDA 334.0 nm
Date Acquired:	2017-3-6 16:33:39 CST		
Date Processed:	2017-3-6 17:04:52 CST		

Auto-Scaled Chromatogram



Peak Results

	RT	Area	% Area	USP Plate Count	USP Resolution
1	13.519	14743884	98.22	40311.96	
2	14.094	266659	1.78	41235.64	2.05