

产品分析证书
Certificate of Analysis

中文名称：人参皂苷RH4

English Name：Ginsenoside Rh4

别名(Alias):

产品编码(Cat. No.):BP0670


CAS Number: 174721-08-5

分子式(M. F.): C₃₆H₆₀O₈

分子量(M. W.): 620.868

批号(Batch No.): PRF8101204

报告日期(Report date): 2017-10-12



检验结果 (Analytical result) :

检验项目 (Test Item)	检验指标 (Specifications)	检验结果 (Results)
外观Appearance	Off-white cotton-like solid	Off-white cotton-like solid
干燥失重Loss on drying	< 3.0%	1.9%
纯度Purity (HPLC-ELSD)*	≥98.0%	99.88%
质谱Mass	620.8±1	Conforms
核磁NMR	Comply with the structure	Conforms

* 色谱图见附件 (Please find HPLC chromatography attached.)

检测方法 (Test Method) : Column: Thermo BDS Hypersil C18 , 4.6*100mm , 2.4um; Column temperature: 30°C; Detection Mode: ELSD;Flow Rate: 1.0ml/min; Sample dissolution: Methanol; Mobile Phase: A, Water B, Acetonitrile; Gradient elution: B, 35%-50%,15min.

贮存条件 (Storage) :2~8°C, protected from light, keep package airtight when not in use.

复测期 (Retest date) :two years (2019-10-11) under conditions list above.

QC: Zhang Ling

Date: 2017-10-12

QA: Wu Qi

Date: 2017-10-12

备注(Remarks): The sample solutions should be prepared within 15 days after receipt. If the day, it is the best preparing the solutions immediately before use. If the solutions have to be made up in advance, it should be made as aliquots in tightly sealed vials at less than -20°C. Generally, these might be useable for up to two weeks.

In case of quality issue, please contact us within 15 days after receipt of the product.


Tel: +86-28-82633397 Fax: +86-28-82633165

http://www.phytopurify.com Email: sales@biopurify.com biopurify@gmail.com

SAMPLE INFORMATION

Sample Name:	Ginsenoside Rh4 PRF8101204	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	
Vial:	99	Acq. Method Set:	Ginsenoside Rh4
Injection #:	1	Processing Method:	Samples
Injection Volume:	10.00 ul	Channel Name:	SATIN
Run Time:	25.0 Minutes	Proc. Chnl. Descr.:	
Date Acquired:	2017-10-12 13:26:15 CST		
Date Processed:	2017-10-12 15:45:27 CST		

Auto-Scaled Chromatogram



Peak Results

	Name	RT	Area	% Area	USP Plate Count	USP Resolution
1		4.733	3657	0.07	93582.1	
2		8.631	2833	0.05	326655.0	63.8
3		11.768	5572018	99.88	49847.7	23.1