

产品分析证书
Certificate of Analysis

中文名称: 阿魏酸异丙酯

English Name: Isopropyl ferulate

别名 (Alias): 4-Hydroxy-3-methoxycinnamic acid isopropyl ester; 3-(4-Hydroxy-3-methoxyphenyl)-2-propenoic acid 1-methylethyl ester

产品编码 (Cat. No.): BP3937


CAS Number: 59831-94-6

分子式 (M. F.): C₁₃H₁₆O₄

分子量 (M. W.): 236.267

批号 (Batch No.): PRF8050521

报告日期 (Report date): 2017-05-05



检验结果 (Analytical result):

检验项目 (Test Item)	检验指标 (Specifications)	检验结果 (Results)
外观 Appearance	Off-white powder	Off-white powder
干燥失重 Loss on drying	<3.0%	1.04%
纯度 Purity (HPLC-DAD, 324nm)*	≥98.0%	99.96%
质谱 Mass	236.2±1	Conforms
核磁 NMR	Comply with the structure	Conforms

* 色谱图见附件 (Please find HPLC chromatography attached.)

检测方法 (Test Method): Column: Thermo BDS Hypersil C18, 4.6*100mm, 2.4μm; Column temperature: 30°C; Detection Mode: UV324nm;

Flow Rate: 0.8ml/min; Sample dissolution: Methanol; Mobile Phase: A, 0.1% Phosphoric acid in water B, Acetonitrile; Gradient elution: B, 25%-45%, 15min.

贮存条件 (Storage): 2~8°C, protected from light, keep package airproofed when not in use.

复测期 (Retest date): two years (2019-05-04) under conditions list above.

QC:

Zhang Ling

Date: 2017-05-05

QA:

Wu Qi

Date: 2017-05-05

备注 (Remarks): The sample solutions should be prepared and used immediately before use. The best preparing the solutions immediately before use. If the solutions have to be made up in advance, it should be made as aliquots in tightly sealed vials at less than -20°C. Generally, these might be useable for up to two weeks.
In case of quality issue, please contact us within 15 days after receipt of the product.


Tel: +86-28-82633397 Fax: +86-28-82633165

http://www.phytopurify.com Email: sales@biopurify.com biopurify@gmail.com

SAMPLE INFORMATION

Sample Name:	Isopropyl ferulate PRF8050521	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	
Vial:	102	Acq. Method Set:	Isopropyl ferulate
Injection #:	1	Processing Method:	Samples
Injection Volume:	5.00 ul	Channel Name:	324.0nm@1
Run Time:	25.0 Minutes	Proc. Chnl. Descr.:	PDA 324.0 nm
Date Acquired:	2017-5-5 13:12:33 CST		
Date Processed:	2017-5-5 14:00:04 CST		

Auto-Scaled Chromatogram



Peak Results

	Name	RT	Area	% Area	USP Plate Count	USP Resolution
1		7.106	3457	0.02	7129.3	
2		10.305	4284	0.02	17198.0	14.1
3		13.640	20488165	99.96	44119.6	14.7